

## 产品质检报告

## Certificate of Analysis (CoA)

产品名称 Product Name:  $\beta$ -内酰胺类检测试纸条  
产品批号 Lot No.: C161240101  
生产日期 Date of Production: 2024.01.21  
有效期至 Expired: 2025.07  
质检编码 QC Code: C161240101

检测项目 Test item	技术要求 Standard	检验结果 Test result	单项判定 Item conclusion
物理外观 Physical Appearance	1、试纸条规格 Dipsticks Specification : 4.50mm×76mm 2、包装洁净平整, 密封严实 Good Packaging 3、内容物齐全 Completed Content 4、试纸条洁净无破损含干燥剂 No damaged Dipsticks. Including Desiccant 5、产品规格: 8T/桶, 96T/盒 Product specification: 8T/Barrel , 96T/Box	符合要求 Conformed	合格 Qualified
检测限 LOD	1、检测浓度 0ng/ml 时, 样品呈阴性 0ng/ml, negative 2、添加检测限浓度, 样品呈阳性 LOD concentration, positive	符合要求 Conformed	合格 Qualified
假阴性率 False Negative Rate	0	符合要求 Conformed	合格 Qualified
假阳性率 False Positive Rate	$\leq 2\%$	符合要求 Conformed	合格 Qualified
均一性 Uniformity	产品显色均一 Uniformed Stripe Color	符合要求 Conformed	合格 Qualified
结论 Conclusion:	本产品经检验合格 Conform of Specification		

质量管理部 QC Department

日期 Date: 2024.01.21

